

# EARL PET/CT | PET/MR

## Accreditation 2024

THIS IS TO CERTIFY THAT THE

### Institute of nuclear technologies for health INTECNUS

Department of Nuclear Medicine

Scanner: GE Discovery 710 PET14244PT8

is granted accreditation according to the requirements and specifications indicated in the EARL accreditation user manual for the following oncology isotope imaging and accreditation program: **<sup>18</sup>F standards 2**

PET/CT | PET/MR scanner accreditation ensures similar performance of the systems within a multicentre setting by harmonising reconstructions for quantification of PET/CT | PET/MR scans. Accredited PET/CT | PET/MR centres of excellence can compare, exchange and combine PET/CT | PET/MR findings, including SUVs, since data are collected and processed in a standardised manner. The regular submission and approval of QC data, as required by EARL, is a condition for the maintenance of the accreditation.

Date of Print: Q1-2024



Rudi A.J.O. Dierckx  
EANM President 2023-2024



Ivalina Hristova  
Programme Director

